



Republic of the Philippines
 Department of Health
 Regional Office I
MARIANO MARCOS MEMORIAL HOSPITAL AND MEDICAL CENTER
 City of Batac, Ilocos Norte
 Trunk line 077-600-8000; Fax line 077-792-3133
 e-mail address: mmmh_doh@yahoo.com
 “PHIC Accredited Health Care Provider”
 “ISO 9001:2015 Certified”
 “PGS Compliant”

Bid Bulletin No.: **2020-19**
 Date: **November 3, 2020**

BIDS AND AWARDS COMMITTEE

BID BULLETIN/SUPPLEMENTAL BID

This first (1st) Bid Bulletin is issued to modify or amend the Bid Documents of the project, **SUPPLY, DELIVERY, INSTALLATION AND COMMISSIONING OF VARIOUS MEDICAL EQUIPMENT (Group B)**. This shall form the integral part of Bid Documents.

The following are the changes in the bidding documents:

PARTICULARS	CHANGES/ADDITIONAL REQUIREMENTS
p. 15 Bid Security 14.2 The Bid and bid security shall be valid until March 11, 2020 (120 calendar days from the Opening of Bids on November 11, 2020).	14.2 The Bid and bid security shall be valid until <u>March 11, 2021</u> (120 calendar days from the Opening of Bids on November 11, 2020).
p. 16 Deadline for Submission of Bids 16.1 The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the IB.	16.1 The Bidders shall submit on the specified date and time and either at its physical address or through online submission as indicated in paragraph 7 of the IB.
p. 27 Special Conditions of the Contract Transportation ... Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure in accordance with GCC Clause Error! Reference source not found..	... Contract the period from when the Goods were first ready for shipment and the actual date of shipment the period of delay will be considered force majeure in accordance with GCC Clause Error! Reference source not found..
p. 31 1 UNIT OR TABLE WITH ELECTRIC HYDRAULIC SYSTEM Accessories: <ul style="list-style-type: none"> Complete accessories for the following positions: lithotomy position, prone position, knee chest position, lateral position and with head pillow Standard arm board with 60mm pad Anesthesia screen L type with clamp 	Accessories: <ul style="list-style-type: none"> Complete accessories for the following positions: lithotomy position, prone position, knee chest position, lateral position and with head pillow Standard arm board with 60mm pad Anesthesia screen L type with clamp



Republic of the Philippines
Department of Health
Regional Office I

MARIANO MARCOS MEMORIAL HOSPITAL AND MEDICAL CENTER

City of Batac, Ilocos Norte

Trunk line 077-600-8000; Fax line 077-792-3133

e-mail address: mmmh_doh@yahoo.com

“PHIC Accredited Health Care Provider”

“ISO 9001:2015 Certified”

“PGS Compliant”

	<ul style="list-style-type: none"> - Radiographic table tops - Cassette tray (390 x 390) - Raised armboard with clamp - Arm shield with pad (160mm) - IVR armboard - Chest and waist support with clamps - Lateral support with clamp - Adjustable body support with clamp - Accessory cart <p>Features:</p> <ul style="list-style-type: none"> - 665mm table top slide stroke - Wide imaging area (1,198mm slide to foot/928mm to head) - With foot switch (pedal function for elevation, Trendelenburg and tilt) - Return to normal level in just one click 	<ul style="list-style-type: none"> - Radiographic table tops - Cassette tray (390 x 390) - Raised armboard with clamp - Arm shield with pad (160mm) - IVR armboard - Chest and waist support with clamps - Lateral support with clamp - Adjustable body support with clamp - Accessory cart <p>Features:</p> <ul style="list-style-type: none"> - 665mm table top slide stroke - Wide imaging area (1,198mm slide to foot/928mm to head) - With foot switch (pedal function for elevation, Trendelenburg and tilt) - Return to normal level in just one click
p. 36 OR 4	2 UNITS ELECTRO CAUTERY MACHINE	<p>IV. Additional Specifications</p> <p>9. Other Additional Accessories</p> <ul style="list-style-type: none"> a) Foot Switching Monopolar Laparoscopic Cord b) Foot Switching Bipolar Cord and Cushing Forceps c) Can be connected to smoke evacuator
p. 37 NC 2	10 UNITS CARDIAC MONITOR	<p>Additional Requirements within the warranty period</p> <ul style="list-style-type: none"> a) Bandage type sensor pulse oximeter b) Infant/Neonatal BP Cuff c) Neonatal cardiac patches
p. 37 NC 3	<p>5 UNITS SYRINGE PUMP</p> <ul style="list-style-type: none"> - Accept all syringes (10cc, 20cc, 30cc, 50cc, 100cc) 	<ul style="list-style-type: none"> - Accept all syringes (1cc, 20cc, 30cc, 50cc, 60cc) - With extension set per unit
p. 38 HC 1	<p>2 UNITS TARGET CONTROLLED INFUSION (TCI) SYRINGE PUMP Infusion</p> <ul style="list-style-type: none"> - Flow rate <ul style="list-style-type: none"> • Flow rate can be limited according to drug name (soft and hard limits) with Agilia Vigilant Drug Lib, IV Medication Safety 	<ul style="list-style-type: none"> - Flow rate <ul style="list-style-type: none"> • Flow rate can be limited according to drug name (soft and hard limits) with Agilia Vigilant Drug Lib, IV Medication Safety



Republic of the Philippines
 Department of Health
 Regional Office I
MARIANO MARCOS MEMORIAL HOSPITAL AND MEDICAL CENTER
 City of Batac, Ilocos Norte
 Trunk line 077-600-8000; Fax line 077-792-3133
 e-mail address: mmmh_doh@yahoo.com
 “PHIC Accredited Health Care Provider”
 “ISO 9001:2015 Certified”
 “PGS Compliant”

System.	System.
<ul style="list-style-type: none"> - TCI Mode Pharmacokinetic Models <ul style="list-style-type: none"> • Marsh & Schnider for Propofol administration on adults. • Minto, Gepts & Scott for Remifentanil, Sufentanil and Alfentanil. Administration in adults. 	<ul style="list-style-type: none"> - TCI Mode Pharmacokinetic Models <ul style="list-style-type: none"> • Marsh & Schnider for Propofol administration on adults. • Minto, Gepts & Scott for Remifentanil, Sufentanil and Alfentanil. Administration in adults. • Kataria & Paedfusor for propofol administration on pediatrics

Other Instructions:

- 1. All bids submitted after the deadline shall not be accepted by the BAC. The BAC shall record in the minutes of bid submission and opening, the bidder's name, its representative and the time the late bid was submitted.**
- 2. The bidder shall indicate/state in the Statement of Compliance, in the Technical Specification, against each of the individual parameters of each Specification the corresponding/equivalent performance parameter of the equipment offered.**
- 3. All other provisions of the bidding documents which are not affected shall remain in effect.**

Furthermore, in accordance to Section 41 of the RA 9184 and its IRR, the Mariano Marcos Memorial Hospital and Medical Center reserves the right to reject documents which do not comply with the requirements, waive any formalities of documents or consider any submission of documents as substantial compliance, reject any and all bids, declare failure of bidding, annul the bidding process, or not award the contract, or if the funds/allotment for the program/project/activity has been withheld or reduced through no fault of the procuring entity without thereby incurring any liability to the affected bidder or bidders.

The BAC assumes no responsibility whatsoever to compensate or indemnify bidders for any expenses incurred in the preparation of the bid.

For guidance and information of all concerned.

(sgd)**LOUIE A. GUERRERO, MD, FPCR**
 BAC Chairperson